

# Wet natural ventilation temperature sensor (Pt100 output)





- ▶ Wet temperature measurement with natural ventilation as required for the calculation of the WBGT index (ISO 7243) in fixed and/or outdoor installations
- ▶ Designed as described in the ISO7726 standard for the "Wet Bulb Temperature"
- ▶ 1/2 litre water tank for more than 30 days water autonomy
- ▶ Well protected by the solar radiation
- ▶ In-house ISO17025 calibration laboratory
- ▶ Accessories for 4÷20 mA and RS485 outputs

Thermometric probe with wet bulb and natural ventilation according to ISO7726 standard. The PT100 element is closed inside a wet cotton sock from a water tank. The water tank has a high reflection factor to avoid water temperature increasing due to radiant sources. This sensor is needed for WBGT (ISO7243) heat stress index assessment.

### Technical Specifications


PN	DMA121	
<b>Wet temperature</b>	Principle	Pt100 DIN-A (Class A EN60751)
	Type	RTD 4 wires
	Measuring range	0÷80°C
	Accuracy	0,15°C (@0°C)
	Output	Pt100 DIN-IEC 751 table (EN 60751)
	Resolution	0,01°C
	Response time (T90 Air)	10 min. (air flow 0,2 m/s)
<b>General Information</b>	Standard	ISO7726
	Protection	IP66
	Power consumption	None
	Operative temperature	0÷80°C
	Cable	L. = 5 m.
	Mounting	On DYA049
	Water tank capacity	500 cc
	Water refill	Every 30 days (25°C, 60% RH)
Data logger compatibility	M-Log (ELO008) R-Log (ELR515) E-Log, A-Log	

## Accessories

	<b>DYA032.1</b>	Arm for fixing DMA121 sensor on DYA049 collar
	<b>DYA049</b>	Mast-mounting collar for Ø 45-65 mm. pipe
	<b>SVICA0103</b>	ISO9001 type calibration certificate
	<b>SVACA0105</b>	ISO17025-ACCREDIA type calibration certificate
	<b>DEA420.1</b>	Converter: Output: 4 ÷ 20 mA. Power supply: 10÷30 Vac/dc More information on MW9008 catalogue
	<b>MDMMA1010.1</b>	Same feature as DEA420.1: Output: RS485 Modbus-RTU



▶ LSI LASTEM is an ISO17025 accredited laboratory for air temperature measurements. All sensors manufactured are tested inside our laboratory. LSI LASTEM provides Test report for any sensor supplied and on request, ISO17025 or ISO9001 calibration certificates (see Accessories list).



▶ Outdoor measurement of the wet temperature finds its most typical application of the calculation of the WBGT index (ISO 7243) when continuous or outdoor monitoring is required. In these cases, small sensors with little water capacity and inadequate solar shields are not suitable. The DMA121 sensor is the "meteorological" solution for this measurement.